
Midmark M11 - v1.0.0
Total Cycles: 963

Name: RITTER 1
Sterilizer ID: V2178198
07/05/2021 02:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 214.5 F 0.0
2:00 212.8 F 4.2
4:00 207.8 F 9.2
6:00 224.5 F 8.6
8:00 232.0 F 9.9
10:00 237.7 F 8.6
12:00 244.6 F 13.3
14:00 243.7 F 12.2
16:00 256.8 F 19.1
18:00 268.4 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.5
1:00 272.8 F 29.2
2:00 271.7 F 28.5
3:00 272.6 F 29.0
4:00 271.7 F 28.6

Min 271.7 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.7 F 0.0
5:00 232.3 F 0.0
10:00 234.1 F 0.0
15:00 222.9 F 0.0
20:00 210.7 F 0.0
25:00 203.0 F 0.0
Dry Cycle Aborted

FILLING: 2:04
HEATING: 19:04
STERILIZING: 4:00
VENTING: 3:14
DRYING: 28:31
TOTAL CYCLE: 00:56:53

07/05/2021 02:57 PM
ACCEPTED: X
REJECTED:
Operator: U01