

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000733  
14:27 22/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:18  
PRESSURIZING 2:31  
STERILIZING 3:24  
136.0 C 228kPa 3:24  
136.4 C 233kPa 3:54  
136.3 C 233kPa 4:24  
136.2 C 230kPa 4:54  
136.3 C 233kPa 5:24  
136.4 C 233kPa 5:54  
136.3 C 232kPa 6:24  
136.3 C 232kPa 6:54  
136.3 C 232kPa 6:54  
VENTING 6:55  
AIR DRYING 7:45  
CYCLE COMPLETE 8:04  
-----

Digital Signature #  
90AB2FC5C7FA2A51  
-----