
Midmark M11 - v1.0.0
Total Cycles: 2322

Name: RITTER 1
Sterilizer ID: V2178198
09/14/2022 07:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.3 F 0.0
2:00 74.6 F 0.0
4:00 83.7 F 0.0
6:00 109.7 F 0.0
8:00 147.3 F 0.1
10:00 179.0 F 0.6
12:00 198.0 F 4.2
14:00 211.9 F 7.2
16:00 222.7 F 10.6
18:00 229.8 F 8.3
20:00 234.2 F 10.0
22:00 236.6 F 9.0
24:00 244.5 F 12.8
26:00 242.5 F 11.4
28:00 255.3 F 18.0
30:00 266.3 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.5 F 28.8
2:00 271.4 F 28.0
3:00 272.4 F 28.9
4:00 272.0 F 28.4

Min 271.3 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 222.9 F 0.0
10:00 228.6 F 0.0
15:00 212.7 F 0.0
20:00 237.0 F 0.0
25:00 243.4 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:40
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:09:36
CYCLE COMPLETE

09/14/2022 08:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____