
Midmark M11 - v1.0.0
Total Cycles: 5328

Name: RITTER 1
Sterilizer ID: V2178198
02/10/2025 01:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.0 F 0.0
2:00 148.8 F 2.5
4:00 166.4 F 5.8
6:00 187.9 F 9.7
8:00 224.6 F 7.4
10:00 219.9 F 7.9
12:00 232.5 F 12.9
14:00 236.4 F 9.2
16:00 238.8 F 9.5
18:00 246.4 F 13.4
20:00 258.3 F 19.8
22:00 268.9 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.3 F 28.9
2:00 273.3 F 29.3
3:00 272.5 F 29.0
4:00 272.3 F 29.2

Min 270.9 F 28.2
Max 273.4 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.2 F 0.0
5:00 230.1 F 0.0
10:00 211.3 F 0.0
15:00 190.8 F 0.0
20:00 213.6 F 0.0
25:00 185.8 F 0.0
30:00 157.2 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 23:19
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:01:51
CYCLE COMPLETE

02/10/2025 02:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____