
Midmark M11 - v1.0.0
Total Cycles: 3787

Name: RITTER 1
Sterilizer ID: V2178198
09/18/2023 10:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.5 F 0.0
2:00 168.4 F 1.8
4:00 187.7 F 4.9
6:00 203.0 F 9.0
8:00 217.9 F 13.4
10:00 222.7 F 7.9
12:00 239.1 F 12.5
14:00 235.5 F 9.3
16:00 239.8 F 10.2
18:00 246.0 F 13.1
20:00 256.8 F 19.1
22:00 267.1 F 25.3
24:00 272.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 272.4 F 28.9
2:00 272.8 F 29.3
3:00 271.6 F 28.8
4:00 272.4 F 29.0

Min 271.5 F 28.6
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.5 F 0.0
5:00 223.7 F 0.0
10:00 227.1 F 0.0
15:00 228.7 F 0.0
20:00 220.5 F 0.0
25:00 212.7 F 0.0
30:00 235.2 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 24:10
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:03:01
CYCLE COMPLETE

09/18/2023 11:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____