
Midmark M11 - v1.0.0
Total Cycles: 603

Name: MIDMARK
Sterilizer ID: V2178198
01/14/2021 07:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.6 F 0.0
2:00 74.1 F 0.0
4:00 88.3 F 0.1
6:00 122.3 F 0.2
8:00 154.6 F 0.3
10:00 184.2 F 2.1
12:00 202.6 F 5.3
14:00 213.9 F 9.3
16:00 226.1 F 13.4
18:00 229.9 F 8.3
20:00 241.2 F 13.2
22:00 239.6 F 10.7
24:00 236.7 F 9.0
26:00 251.8 F 16.1
28:00 263.9 F 23.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.1 F 28.4
2:00 272.7 F 28.9
3:00 271.6 F 28.2
4:00 272.3 F 28.5

Min 271.5 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.7 F 0.0
5:00 199.7 F 0.0
10:00 197.9 F 0.0
15:00 195.2 F 0.0
20:00 194.8 F 0.0
25:00 225.8 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 29:57
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:07:58
CYCLE COMPLETE

01/14/2021 09:00 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____