

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3058

Name: RITTER2  
Sterilizer ID: V2313228  
10/27/2023 12:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.7 F 0.0  
2:00 154.9 F 2.4  
4:00 171.6 F 6.2  
6:00 198.6 F 11.1  
8:00 215.2 F 5.4  
10:00 228.4 F 10.8  
12:00 229.8 F 7.8  
14:00 240.6 F 11.6  
16:00 245.2 F 13.4  
18:00 258.1 F 20.4  
20:00 269.7 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.1  
1:00 272.6 F 29.9  
2:00 271.7 F 29.1  
3:00 272.0 F 29.9  
4:00 272.4 F 29.6  
  
Min 271.3 F 29.0  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.9 F 0.0  
5:00 213.2 F 0.0  
10:00 212.3 F 0.0  
15:00 207.3 F 0.0  
20:00 205.3 F 0.0  
25:00 203.8 F 0.0  
30:00 205.3 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 20:59  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 00:59:38  
CYCLE COMPLETE

10/27/2023 01:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_