

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000071  
8:32 12/02/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:44  
PRESSURIZING 3:44  
STERILIZING 5:09  
136.5 C 220kPa 5:09  
136.0 C 221kPa 5:39  
136.4 C 223kPa 6:09  
136.2 C 221kPa 6:39  
136.2 C 222kPa 7:09  
136.4 C 222kPa 7:39  
136.2 C 220kPa 8:09  
136.4 C 224kPa 8:39  
136.4 C 224kPa 8:39  
VENTING 8:39  
AIR DRYING 9:33  
CYCLE COMPLETE 9:40

-----  
Digital Signature #  
8F13A9AE787E0F6E  
-----