

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4774

Name: RITTER 1  
Sterilizer ID: V2178198  
07/29/2024 10:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 137.2 F 0.0  
2:00 136.4 F 1.7  
4:00 153.2 F 4.5  
6:00 178.6 F 7.9  
8:00 213.2 F 11.7  
10:00 216.0 F 3.3  
12:00 221.8 F 7.0  
14:00 232.0 F 11.2  
16:00 229.6 F 6.7  
18:00 242.7 F 12.2  
20:00 239.4 F 10.0  
22:00 251.5 F 15.9  
24:00 262.0 F 21.9  
26:00 271.9 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.5 F 28.9  
2:00 272.1 F 28.3  
3:00 271.2 F 28.2  
4:00 272.5 F 28.9

Min 271.2 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.7 F 0.0  
5:00 219.7 F 0.0  
10:00 227.1 F 0.0  
15:00 224.5 F 0.0  
20:00 217.4 F 0.0  
25:00 215.1 F 0.0  
30:00 211.7 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 26:37  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:05:09  
CYCLE COMPLETE

07/29/2024 11:34 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_