
Midmark M11 - v1.0.5
Total Cycles: 1385

Name: RITTER2
Sterilizer ID: V2313228
08/09/2022 01:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.1 F 0.0
2:00 181.9 F 1.1
4:00 194.7 F 6.1
6:00 213.6 F 11.4
8:00 226.0 F 7.4
10:00 238.0 F 13.2
12:00 239.9 F 11.5
14:00 239.8 F 10.7
16:00 253.4 F 17.7
18:00 265.4 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.9 F 29.7
2:00 271.6 F 29.0
3:00 272.8 F 29.9
4:00 272.4 F 29.4

Min 271.5 F 29.0
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.1 F 0.0
5:00 215.6 F 0.0
10:00 233.1 F 0.0
15:00 227.8 F 0.0
20:00 225.8 F 0.0
25:00 225.3 F 0.0
30:00 225.0 F 0.0

Drying Complete

FILLING: 1:32
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:28
CYCLE COMPLETE

08/09/2022 02:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____