
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 004256
13:39 29/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:32
PRESSURIZING 3:32
STERILIZING 4:35
136.4 C 224kPa 4:35
136.3 C 224kPa 5:05
136.4 C 224kPa 5:35
136.4 C 225kPa 6:05
136.4 C 224kPa 6:35
136.2 C 224kPa 7:05
136.4 C 223kPa 7:35
136.4 C 224kPa 8:05
136.4 C 224kPa 8:05
VENTING 8:05
AIR DRYING 8:56
CYCLE COMPLETE 9:05

Digital Signature #
CC3E1D5DC78F7E02
