
Midmark M11 - v1.0.5
Total Cycles: 2118

Name: RITTER2
Sterilizer ID: V2313228
03/06/2023 08:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.6 F 0.0
2:00 80.5 F 0.0
4:00 106.8 F 0.0
6:00 148.8 F 0.0
8:00 169.4 F 1.4
10:00 208.3 F 4.6
12:00 222.4 F 8.2
14:00 231.5 F 12.0
16:00 227.1 F 6.6
18:00 234.4 F 10.8
20:00 232.2 F 7.3
22:00 239.5 F 12.2
24:00 238.1 F 10.0
26:00 250.5 F 16.1
28:00 261.3 F 22.4
30:00 271.6 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.2 F 29.6
2:00 271.5 F 29.4
3:00 272.1 F 29.5
4:00 271.5 F 29.4

Min 271.2 F 29.0
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.2 F 0.0
5:00 221.3 F 0.0
10:00 224.8 F 0.0
15:00 223.2 F 0.0
20:00 219.8 F 0.0
25:00 220.1 F 0.0
30:00 218.3 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:36
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:08:36
CYCLE COMPLETE

03/06/2023 09:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____