
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000207
16:01 21/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:29
STERILIZING 5:10
136.4 C 223kPa 5:10
136.4 C 233kPa 5:40
136.4 C 233kPa 6:10
136.2 C 232kPa 6:40
136.4 C 233kPa 7:10
136.2 C 232kPa 7:40
136.2 C 231kPa 8:10
136.3 C 232kPa 8:40
136.3 C 232kPa 8:40
VENTING 8:40
AIR DRYING 9:42
CYCLE COMPLETE 10:58

Digital Signature #
7EB3D4AB79868775
