
Midmark M11 - v1.0.0
Total Cycles: 5298

Name: RITTER 1
Sterilizer ID: V2178198
01/30/2025 11:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.6 F 0.0
2:00 170.1 F 2.0
4:00 171.5 F 5.3
6:00 185.7 F 9.0
8:00 203.0 F 13.0
10:00 214.7 F 4.6
12:00 226.5 F 8.7
14:00 237.1 F 13.0
16:00 236.0 F 9.3
18:00 242.1 F 10.7
20:00 245.1 F 12.6
22:00 256.0 F 18.4
24:00 265.8 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 271.7 F 28.8
2:00 272.4 F 29.0
3:00 273.1 F 29.4
4:00 272.4 F 29.0

Min 270.9 F 28.1
Max 273.1 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.4 F 0.0
5:00 216.0 F 0.0
10:00 221.2 F 0.0
15:00 219.4 F 0.0
20:00 223.0 F 0.0
25:00 230.5 F 0.0
30:00 226.9 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 26:00
STERILIZING: 4:00
VENTING: 3:56
DRYING: 30:00
TOTAL CYCLE: 01:04:50
CYCLE COMPLETE

01/30/2025 12:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____