
Midmark M11 - v1.0.5
Total Cycles: 3648

Name: RITTER2
Sterilizer ID: V2313228
05/27/2024 12:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.2 F 0.0
2:00 94.1 F 0.1
4:00 102.9 F 0.9
6:00 134.6 F 3.9
8:00 172.2 F 6.8
10:00 194.0 F 10.1
12:00 209.5 F 8.6
14:00 222.1 F 7.4
16:00 233.9 F 11.4
18:00 230.9 F 7.3
20:00 240.2 F 12.6
22:00 238.7 F 10.7
24:00 250.5 F 16.5
26:00 261.0 F 22.4
28:00 270.8 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.5
1:00 272.7 F 30.2
2:00 271.5 F 29.4
3:00 271.8 F 30.1
4:00 272.4 F 30.0

Min 271.0 F 29.3
Max 272.9 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 203.8 F 0.0
10:00 196.0 F 0.0
15:00 186.5 F 0.0
20:00 195.2 F 0.0
25:00 197.9 F 0.0
30:00 202.3 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 28:50
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:07:19
CYCLE COMPLETE

05/27/2024 02:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____