

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4479

Name: RITTER 1  
Sterilizer ID: V2178198  
04/23/2024 08:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.2 F 0.0  
2:00 83.1 F 0.3  
4:00 82.9 F 1.3  
6:00 107.3 F 3.4  
8:00 141.9 F 5.9  
10:00 172.9 F 9.0  
12:00 201.9 F 12.8  
14:00 212.4 F 4.0  
16:00 222.6 F 7.4  
18:00 233.4 F 11.1  
20:00 228.6 F 6.3  
22:00 239.7 F 11.1  
24:00 234.6 F 8.2  
26:00 248.5 F 14.5  
28:00 259.4 F 20.5  
30:00 269.6 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.1 F 28.5  
2:00 272.4 F 29.0  
3:00 271.4 F 28.4  
4:00 271.9 F 28.7

Min 271.3 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.4 F 0.0  
5:00 223.6 F 0.0  
10:00 209.3 F 0.0  
15:00 193.6 F 0.0  
20:00 197.0 F 0.0  
25:00 184.8 F 0.0  
30:00 183.1 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:60  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:09:10  
CYCLE COMPLETE

04/23/2024 09:19 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_