
Midmark M11 - v1.0.0
Total Cycles: 2251

Name: RITTER 1
Sterilizer ID: V2178198
08/25/2022 03:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.7 F 0.0
2:00 188.4 F 0.5
4:00 204.5 F 3.3
6:00 222.1 F 8.6
8:00 234.2 F 10.6
10:00 239.7 F 11.6
12:00 238.8 F 10.1
14:00 237.5 F 9.2
16:00 253.4 F 16.8
18:00 265.3 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.4 F 28.5
2:00 272.1 F 28.7
3:00 272.0 F 28.3
4:00 272.4 F 28.9

Min 271.4 F 28.0
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.6 F 0.0
5:00 230.0 F 0.0
10:00 248.6 F 0.0
15:00 241.1 F 0.0
20:00 237.1 F 0.0
25:00 228.2 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 19:43
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:25
CYCLE COMPLETE

08/25/2022 04:14 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____