

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3697

Name: RITTER 1  
Sterilizer ID: V2178198  
08/29/2023 11:17 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.5 F 0.0  
2:00 177.3 F 3.5  
4:00 199.5 F 8.1  
6:00 216.2 F 12.3  
8:00 226.6 F 8.2  
10:00 237.3 F 13.4  
12:00 238.7 F 10.8  
14:00 236.6 F 9.0  
16:00 250.9 F 15.7  
18:00 262.4 F 22.3  
20:00 272.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.2 F 28.7  
2:00 272.0 F 28.9  
3:00 271.9 F 28.4  
4:00 272.5 F 29.1  
  
Min 271.4 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.8 F 0.0  
5:00 215.9 F 0.0  
10:00 214.7 F 0.0  
15:00 215.7 F 0.0  
20:00 211.7 F 0.0  
25:00 207.7 F 0.0  
30:00 205.9 F 0.0  
Drying Complete

FILLING: 2:24  
HEATING: 20:20  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:00:08  
CYCLE COMPLETE

08/29/2023 12:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_