
Midmark M11 - v1.0.0
Total Cycles: 1723

Name: RITTER 1
Sterilizer ID: V2178198
03/16/2022 12:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.9 F 0.0
2:00 152.4 F 1.4
4:00 167.2 F 4.4
6:00 203.2 F 8.8
8:00 222.3 F 7.8
10:00 227.8 F 8.6
12:00 238.9 F 12.7
14:00 238.3 F 10.5
16:00 235.7 F 8.3
18:00 249.8 F 15.2
20:00 261.9 F 22.0
22:00 272.5 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.0 F 28.7
2:00 272.4 F 29.2
3:00 272.2 F 29.3
4:00 271.7 F 28.6

Min 271.7 F 28.6
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 214.6 F 0.0
10:00 233.6 F 0.0
15:00 230.9 F 0.0
20:00 228.8 F 0.0
25:00 233.0 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 22:24
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:00:37
CYCLE COMPLETE

03/16/2022 01:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____