
Midmark M11 - v1.0.5
Total Cycles: 2675

Name: RITTER2
Sterilizer ID: V2313228
07/20/2023 11:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.9 F 0.0
2:00 163.0 F 2.1
4:00 168.6 F 6.0
6:00 193.6 F 10.8
8:00 214.3 F 4.8
10:00 226.0 F 10.2
12:00 230.5 F 7.9
14:00 242.7 F 13.2
16:00 242.7 F 12.1
18:00 255.2 F 18.7
20:00 266.1 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.0 F 29.3
2:00 272.2 F 29.7
3:00 272.0 F 29.9
4:00 271.6 F 29.2

Min 271.4 F 29.2
Max 272.3 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.4 F 0.0
5:00 231.8 F 0.0
10:00 224.6 F 0.0
15:00 220.8 F 0.0
20:00 221.9 F 0.0
25:00 222.9 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 21:34
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:51
CYCLE COMPLETE

07/20/2023 12:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____