

-----  
Midmark M11 - v1.0.5  
Total Cycles: 562

Name: RITTER2  
Sterilizer ID: V2313228  
11/18/2021 07:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.7 F 0.0  
2:00 74.0 F 0.0  
4:00 78.5 F 0.0  
6:00 111.0 F 0.1  
8:00 148.4 F 0.2  
10:00 183.0 F 0.4  
12:00 213.5 F 4.0  
14:00 220.7 F 7.4  
16:00 230.0 F 11.1  
18:00 227.9 F 6.9  
20:00 241.3 F 12.1  
22:00 238.6 F 10.2  
24:00 236.5 F 8.7  
26:00 250.4 F 15.7  
28:00 261.5 F 22.1  
30:00 271.8 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 272.9 F 29.5  
2:00 271.8 F 28.6  
3:00 272.1 F 29.3  
4:00 272.5 F 29.2

Min 271.4 F 28.5  
Max 272.9 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.9 F 0.0  
5:00 221.6 F 0.0  
10:00 207.2 F 0.0  
15:00 190.2 F 0.0  
20:00 176.6 F 0.0  
25:00 172.2 F 0.0  
30:00 168.1 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 30:36  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:08:44  
CYCLE COMPLETE

11/18/2021 08:55 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_