
Midmark M11 - v1.0.0
Total Cycles: 871

Name: RITTER 1
Sterilizer ID: V2178198
06/08/2021 09:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.7 F 0.0
2:00 146.0 F 1.5
4:00 162.7 F 4.0
6:00 184.7 F 7.5
8:00 204.1 F 11.8
10:00 215.1 F 4.6
12:00 226.9 F 9.0
14:00 238.2 F 13.6
16:00 238.4 F 10.5
18:00 236.4 F 8.6
20:00 248.7 F 14.6
22:00 260.2 F 21.0
24:00 270.7 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 29.1
2:00 272.3 F 28.6
3:00 271.5 F 28.2
4:00 272.4 F 29.1

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.4 F 0.0
5:00 210.4 F 0.0
10:00 228.9 F 0.0
15:00 221.4 F 0.0
20:00 230.7 F 0.0
25:00 229.6 F 0.0
30:00 213.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 24:51
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:03:19
CYCLE COMPLETE

06/08/2021 10:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____