
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005021
9:37 31/01/2023

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 3:17
STERILIZING 4:20
136.0 C 220kPa 4:20
136.0 C 220kPa 4:50
136.4 C 222kPa 5:20
136.0 C 220kPa 5:50
136.2 C 222kPa 6:20
136.4 C 224kPa 6:50
136.4 C 222kPa 7:20
135.9 C 219kPa 7:50
135.9 C 219kPa 7:50
VENTING 7:51
AIR DRYING 8:48
CYCLE COMPLETE 8:51

Digital Signature #
90E734EBE50C5055
