

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1102

Name: RITTER2  
Sterilizer ID: V2313228  
05/12/2022 09:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.5 F 0.0  
2:00 157.7 F 2.1  
4:00 163.4 F 5.8  
6:00 189.2 F 10.9  
8:00 214.9 F 5.3  
10:00 228.9 F 10.7  
12:00 232.5 F 7.1  
14:00 242.5 F 13.2  
16:00 242.0 F 11.9  
18:00 254.6 F 18.3  
20:00 265.6 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 272.2 F 29.9  
2:00 272.3 F 29.5  
3:00 271.5 F 29.0  
4:00 271.2 F 29.1

Min 271.1 F 28.8  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.2 F 0.0  
5:00 198.8 F 0.0  
10:00 208.8 F 0.0  
15:00 213.4 F 0.0  
20:00 214.9 F 0.0  
25:00 207.4 F 0.0  
30:00 203.7 F 0.0  
Drying Complete

FILLING: 1:25  
HEATING: 21:48  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:00:22  
CYCLE COMPLETE

05/12/2022 10:27 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02