
Midmark M11 - v1.0.0
Total Cycles: 3049

Name: RITTER 1
Sterilizer ID: V2178198
03/28/2023 05:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.8 F 0.0
2:00 131.5 F 0.1
4:00 151.9 F 2.8
6:00 182.6 F 6.2
8:00 201.5 F 10.4
10:00 215.0 F 5.7
12:00 223.5 F 8.0
14:00 236.9 F 12.4
16:00 234.6 F 8.9
18:00 241.9 F 10.8
20:00 245.5 F 12.9
22:00 257.3 F 19.2
24:00 268.0 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.4 F 29.0
2:00 272.0 F 28.4
3:00 271.4 F 28.4
4:00 272.4 F 29.0

Min 271.2 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 214.3 F 0.0
10:00 217.2 F 0.0
15:00 216.4 F 0.0
20:00 213.1 F 0.0
25:00 210.8 F 0.0
30:00 209.7 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 25:22
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:03:48
CYCLE COMPLETE

03/28/2023 06:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____