
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000840
15:23 05/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:29
STERILIZING 3:32
136.0 C 220kPa 3:32
136.0 C 231kPa 4:02
136.1 C 231kPa 4:32
136.3 C 232kPa 5:02
136.4 C 233kPa 5:32
136.2 C 230kPa 6:02
136.2 C 231kPa 6:32
135.9 C 229kPa 7:02
135.9 C 229kPa 7:02
VENTING 7:02
AIR DRYING 7:54
CYCLE COMPLETE 18:46

Digital Signature #
99CCE21437217AA9
