
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
2.1uS / 1.3ppm
CYCLE NUMBER 000032
13:54 10/08/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 3:56
STERILIZING 5:00
136.4 C 227kPa 5:00
136.4 C 226kPa 5:30
136.3 C 224kPa 6:00
136.5 C 227kPa 6:30
136.4 C 226kPa 7:00
136.0 C 224kPa 7:30
136.2 C 225kPa 8:00
136.2 C 224kPa 8:30
VENTING 8:30
AIR DRYING 9:22
CYCLE COMPLETE 39:22

Digital Signature #
88CF71EDA41B2859
