

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2973

Name: RITTER2  
Sterilizer ID: V2313228  
10/04/2023 07:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 159.0 F 0.0  
2:00 161.3 F 2.5  
4:00 188.0 F 6.7  
6:00 206.6 F 11.6  
8:00 219.7 F 6.7  
10:00 234.0 F 12.0  
12:00 233.8 F 9.5  
14:00 238.0 F 9.5  
16:00 247.7 F 14.6  
18:00 259.9 F 21.3  
20:00 270.2 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.5  
1:00 271.8 F 29.4  
2:00 272.6 F 29.7  
3:00 272.3 F 30.0  
4:00 271.4 F 29.8  
  
Min 271.2 F 29.4  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.0 F 0.0  
5:00 238.1 F 0.0  
10:00 190.3 F 0.0  
15:00 181.4 F 0.0  
20:00 177.6 F 0.0  
25:00 174.5 F 0.0  
30:00 172.0 F 0.0  
Drying Complete

FILLING: 1:38  
HEATING: 21:40  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:00:49  
CYCLE COMPLETE

10/04/2023 08:12 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_