
Midmark M11 - v1.0.5
Total Cycles: 3163

Name: RITTER2
Sterilizer ID: V2313228
11/27/2023 02:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.5 F 0.0
2:00 147.3 F 1.7
4:00 170.9 F 5.0
6:00 191.0 F 9.7
8:00 218.1 F 7.2
10:00 222.9 F 9.7
12:00 232.5 F 8.9
14:00 241.2 F 12.4
16:00 240.4 F 11.1
18:00 253.6 F 17.7
20:00 264.9 F 24.5
22:00 271.1 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 272.6 F 29.7
1:00 272.8 F 30.1
2:00 272.2 F 30.3
3:00 271.6 F 29.6
4:00 272.4 F 30.1

Min 271.5 F 29.6
Max 273.2 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.9 F 0.0
5:00 237.4 F 0.0
10:00 217.5 F 0.0
15:00 187.5 F 0.0
20:00 182.0 F 0.0
25:00 181.2 F 0.0
30:00 180.1 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 22:28
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:00:59
CYCLE COMPLETE

11/27/2023 03:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____