
Midmark M11 - v1.0.0
Total Cycles: 4331

Name: RITTER 1
Sterilizer ID: V2178198
02/26/2024 09:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.8 F 0.0
2:00 69.2 F 0.2
4:00 80.5 F 1.1
6:00 109.4 F 3.0
8:00 136.6 F 5.4
10:00 162.0 F 8.3
12:00 180.6 F 11.4
14:00 202.5 F 2.9
16:00 210.1 F 5.3
18:00 219.9 F 8.4
20:00 229.6 F 12.0
22:00 229.2 F 7.2
24:00 240.2 F 11.9
26:00 237.6 F 9.4
28:00 249.4 F 15.1
30:00 259.6 F 20.7
32:00 269.2 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.5 F 29.2
2:00 271.6 F 28.5
3:00 271.7 F 28.9
4:00 272.3 F 29.0

Min 271.1 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.4 F 0.0
5:00 212.7 F 0.0
10:00 212.9 F 0.0
15:00 229.5 F 0.0
20:00 229.1 F 0.0
25:00 230.8 F 0.0
30:00 227.8 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 33:11
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:11:41
CYCLE COMPLETE

02/26/2024 10:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____