

-----  
Midmark M11 - v1.0.0  
Total Cycles: 604

Name: MIDMARK  
Sterilizer ID: V2178198  
01/14/2021 10:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 98.5 F 0.0  
2:00 103.8 F 0.0  
4:00 132.0 F 0.1  
6:00 161.5 F 0.3  
8:00 186.1 F 3.0  
10:00 203.8 F 7.2  
12:00 218.4 F 11.4  
14:00 222.6 F 5.8  
16:00 234.2 F 11.0  
18:00 233.4 F 8.0  
20:00 243.2 F 11.8  
22:00 245.3 F 12.8  
24:00 257.9 F 19.6  
26:00 269.1 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.5 F 28.6  
2:00 272.3 F 28.9  
3:00 271.9 F 28.3  
4:00 272.8 F 29.0

Min 271.4 F 28.2  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.1 F 0.0  
5:00 200.6 F 0.0  
10:00 197.8 F 0.0  
15:00 196.4 F 0.0  
20:00 194.4 F 0.0  
25:00 193.0 F 0.0  
30:00 197.7 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 27:03  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:05:11  
CYCLE COMPLETE

01/14/2021 11:59 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_