
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000953
11:41 25/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:51
PRESSURIZING 3:53
STERILIZING 5:05
136.4 C 232kPa 5:05
136.2 C 231kPa 5:35
136.2 C 231kPa 6:05
136.3 C 232kPa 6:35
136.4 C 232kPa 7:05
136.3 C 233kPa 7:35
136.3 C 231kPa 8:05
136.2 C 231kPa 8:35
136.2 C 231kPa 8:35
VENTING 8:35
AIR DRYING 9:30
CYCLE COMPLETE 10:39

Digital Signature #
287C916EABBD265F
