
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000731
12:43 22/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:35
PRESSURIZING 2:41
STERILIZING 3:42
136.4 C 233kPa 3:42
136.1 C 231kPa 4:12
136.4 C 233kPa 4:42
136.1 C 231kPa 5:12
136.2 C 230kPa 5:42
136.3 C 232kPa 6:12
136.0 C 231kPa 6:42
136.4 C 231kPa 7:12
136.4 C 231kPa 7:12
VENTING 7:12
AIR DRYING 8:03
CYCLE COMPLETE 8:06

Digital Signature #
CC697374700FFA4D
