
Midmark M11 - v1.0.0
Total Cycles: 3081

Name: RITTER 1
Sterilizer ID: V2178198
04/10/2023 11:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.9 F 0.0
2:00 184.7 F 3.2
4:00 207.4 F 8.3
6:00 224.8 F 9.9
8:00 231.9 F 9.6
10:00 236.2 F 8.7
12:00 245.1 F 13.2
14:00 244.3 F 12.1
16:00 257.4 F 19.1
18:00 269.1 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.0
1:00 272.4 F 28.9
2:00 272.4 F 28.6
3:00 271.7 F 28.1
4:00 272.3 F 29.0

Min 271.0 F 27.8
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.1 F 0.0
5:00 236.4 F 0.0
10:00 219.9 F 0.0
15:00 220.1 F 0.0
20:00 219.8 F 0.0
25:00 219.3 F 0.0
30:00 218.3 F 0.0
Drying Complete

FILLING: 2:00
HEATING: 19:09
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:21
CYCLE COMPLETE

04/10/2023 12:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____