

-----  
Midmark M11 - v1.0.0  
Total Cycles: 647

Name: MIDMARK  
Sterilizer ID: V2178198  
02/01/2021 02:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.5 F 0.0  
2:00 173.9 F 3.4  
4:00 189.9 F 8.4  
6:00 235.3 F 8.0  
8:00 230.6 F 10.7  
10:00 234.4 F 7.9  
12:00 239.8 F 10.0  
14:00 248.0 F 13.8  
16:00 261.3 F 21.2  
18:00 272.6 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 27.9  
1:00 272.1 F 28.1  
2:00 272.2 F 28.6  
3:00 272.2 F 28.3  
4:00 272.6 F 28.5  
  
Min 271.0 F 27.7  
Max 273.1 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.2 F 0.0  
5:00 242.5 F 0.0  
10:00 231.5 F 0.0  
15:00 228.4 F 0.0  
20:00 227.5 F 0.0  
25:00 226.8 F 0.0  
30:00 224.0 F 0.0  
  
Drying Complete

FILLING: 1:19  
HEATING: 18:16  
STERILIZING: 4:00  
VENTING: 2:52  
DRYING: 30:00  
TOTAL CYCLE: 00:56:27  
CYCLE COMPLETE

02/01/2021 03:27 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01