
Midmark M11 - v1.0.0
Total Cycles: 3390

Name: RITTER 1
Sterilizer ID: V2178198
06/14/2023 01:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.9 F 0.0
2:00 174.1 F 1.1
4:00 199.8 F 5.7
6:00 217.7 F 11.0
8:00 225.1 F 6.6
10:00 238.1 F 12.1
12:00 236.7 F 9.8
14:00 237.9 F 8.7
16:00 249.9 F 15.4
18:00 262.2 F 22.4
20:00 272.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.5
1:00 272.5 F 29.1
2:00 271.6 F 28.7
3:00 272.2 F 28.8
4:00 272.0 F 29.1

Min 271.4 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 213.5 F 0.0
10:00 240.1 F 0.0
15:00 239.6 F 0.0
20:00 229.5 F 0.0
25:00 219.7 F 0.0
30:00 215.1 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 20:07
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:58:21
CYCLE COMPLETE

06/14/2023 02:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____