
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 003649
12:20 11/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:40
STERILIZING 3:47
136.4 C 220kPa 3:47
136.3 C 224kPa 4:17
136.4 C 225kPa 4:47
136.4 C 224kPa 5:17
136.1 C 224kPa 5:47
136.3 C 224kPa 6:17
136.4 C 226kPa 6:47
136.4 C 224kPa 7:17
136.4 C 224kPa 7:17
VENTING 7:18
AIR DRYING 8:07
CYCLE COMPLETE 8:22

Digital Signature #
9CD62354A733C0B8
