
Midmark M11 - v1.0.0
Total Cycles: 3765

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2023 11:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 119.5 F 0.0
2:00 120.6 F 0.0
4:00 139.9 F 1.0
6:00 165.7 F 3.3
8:00 194.3 F 6.9
10:00 215.1 F 11.5
12:00 224.4 F 5.0
14:00 231.8 F 9.5
16:00 237.6 F 10.4
18:00 240.5 F 10.8
20:00 235.5 F 8.4
22:00 249.6 F 14.8
24:00 260.7 F 21.0
26:00 271.0 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.3 F 28.5
2:00 271.6 F 28.4
3:00 272.1 F 28.5
4:00 271.3 F 28.2

Min 271.2 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.3 F 0.0
5:00 233.6 F 0.0
10:00 211.7 F 0.0
15:00 203.3 F 0.0
20:00 227.3 F 0.0
25:00 230.9 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 26:42
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:04:53
CYCLE COMPLETE

09/13/2023 12:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____