
Midmark M11 - v1.0.0
Total Cycles: 3066

Name: RITTER 1
Sterilizer ID: V2178198
04/03/2023 10:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.5 F 0.0
2:00 140.2 F 0.0
4:00 165.2 F 1.4
6:00 191.8 F 4.6
8:00 219.5 F 8.9
10:00 226.8 F 9.5
12:00 233.1 F 8.6
14:00 241.6 F 12.1
16:00 239.4 F 10.9
18:00 236.1 F 8.9
20:00 250.8 F 15.7
22:00 262.7 F 22.6
24:00 272.0 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 271.9 F 28.7
2:00 272.4 F 29.2
3:00 271.4 F 28.4
4:00 272.3 F 29.0

Min 271.3 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.0 F 0.0
5:00 222.3 F 0.0
10:00 198.7 F 0.0
15:00 209.3 F 0.0
20:00 210.1 F 0.0
25:00 209.6 F 0.0
30:00 210.1 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 24:08
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:02:10
CYCLE COMPLETE

04/03/2023 11:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____