
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007847
8:53 10/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:29
PRESSURIZING 1:33
STERILIZING 2:33
136.0 C 220kPa 2:33
136.1 C 222kPa 3:03
136.1 C 218kPa 3:33
136.4 C 223kPa 4:03
136.1 C 219kPa 4:33
136.3 C 223kPa 5:03
135.8 C 218kPa 5:33
136.4 C 223kPa 6:03
136.4 C 223kPa 6:03
VENTING 6:04
AIR DRYING 6:53
CYCLE COMPLETE 7:16

Digital Signature #
737C9A21ED1FCE6C
