

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 007262  
6:36 25/04/2024

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:02  
PRESSURIZING 3:05  
STERILIZING 4:04  
136.4 C 221kPa 4:04  
136.4 C 223kPa 4:34  
136.1 C 222kPa 5:04  
136.2 C 220kPa 5:34  
136.3 C 222kPa 6:04  
136.4 C 221kPa 6:34  
136.3 C 220kPa 7:04  
136.1 C 221kPa 7:34  
136.1 C 221kPa 7:34  
VENTING 7:34  
AIR DRYING 8:31  
CYCLE COMPLETE 8:51  
-----

Digital Signature #  
3A0371E7131E596D  
-----