
Midmark M11 - v1.0.0
Total Cycles: 3044

Name: RITTER 1
Sterilizer ID: V2178198
03/28/2023 10:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.8 F 0.0
2:00 159.9 F 2.2
4:00 177.0 F 5.7
6:00 198.4 F 10.2
8:00 219.4 F 5.3
10:00 226.9 F 9.6
12:00 237.0 F 9.2
14:00 241.3 F 11.5
16:00 238.7 F 9.6
18:00 252.2 F 16.2
20:00 263.6 F 22.8
22:00 271.8 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.2 F 28.7
2:00 271.1 F 28.1
3:00 272.2 F 28.7
4:00 272.5 F 28.7

Min 271.0 F 28.0
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.6 F 0.0
5:00 218.5 F 0.0
10:00 230.3 F 0.0
15:00 231.0 F 0.0
20:00 229.9 F 0.0
25:00 226.8 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 22:06
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:10
CYCLE COMPLETE

03/28/2023 11:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____