

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1441

Name: RITTER 1  
Sterilizer ID: V2178198  
12/09/2021 11:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.0 F 0.0  
2:00 193.2 F 3.3  
4:00 202.1 F 8.0  
6:00 221.3 F 13.5  
8:00 229.2 F 8.2  
10:00 241.2 F 13.5  
12:00 240.9 F 11.2  
14:00 238.6 F 9.6  
16:00 253.2 F 16.8  
18:00 265.2 F 23.9

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.0  
1:00 272.6 F 28.5  
2:00 272.0 F 28.6  
3:00 271.8 F 27.9  
4:00 272.5 F 28.6  
  
Min 271.4 F 27.9  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.4 F 0.0  
5:00 214.5 F 0.0  
10:00 201.2 F 0.0  
15:00 197.4 F 0.0  
20:00 193.2 F 0.0  
25:00 188.4 F 0.0  
30:00 184.3 F 0.0  
Drying Complete

FILLING: 2:23  
HEATING: 19:32  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:59:01  
CYCLE COMPLETE

12/09/2021 12:34 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02