

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3644

Name: RITTER 1  
Sterilizer ID: V2178198  
08/14/2023 12:33 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.1 F 0.0  
2:00 167.6 F 2.9  
4:00 187.3 F 6.7  
6:00 207.4 F 11.2  
8:00 219.4 F 5.0  
10:00 234.2 F 10.0  
12:00 234.5 F 8.9  
14:00 242.9 F 12.3  
16:00 240.8 F 10.7  
18:00 253.2 F 16.8  
20:00 263.9 F 23.1  
22:00 271.9 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 271.8 F 28.3  
2:00 272.0 F 28.5  
3:00 272.4 F 28.9  
4:00 271.6 F 28.7

Min 271.1 F 28.2  
Max 272.4 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.2 F 0.0  
5:00 228.1 F 0.0  
10:00 221.6 F 0.0  
15:00 221.1 F 0.0  
20:00 220.0 F 0.0  
25:00 224.5 F 0.0  
30:00 227.0 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 22:05  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:01:22  
CYCLE COMPLETE

08/14/2023 01:34 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_