
Midmark M11 - v1.0.5
Total Cycles: 2478

Name: RITTER2
Sterilizer ID: V2313228
06/07/2023 10:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.3 F 0.0
2:00 202.9 F 3.5
4:00 200.0 F 8.9
6:00 222.6 F 8.9
8:00 230.0 F 10.4
10:00 232.1 F 7.8
12:00 241.1 F 11.3
14:00 245.5 F 13.6
16:00 258.2 F 20.5
18:00 269.2 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.4 F 29.6
2:00 271.7 F 29.6
3:00 272.4 F 29.5
4:00 271.6 F 29.5

Min 271.4 F 29.1
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 230.3 F 0.0
10:00 234.0 F 0.0
15:00 230.1 F 0.0
20:00 227.1 F 0.0
25:00 226.0 F 0.0
30:00 224.7 F 0.0

Drying Complete

FILLING: 2:11
HEATING: 19:03
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:26
CYCLE COMPLETE

06/07/2023 11:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____