
Midmark M11 - v1.0.5
Total Cycles: 2869

Name: RITTER2
Sterilizer ID: V2313228
09/12/2023 11:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.4 F 0.0
2:00 163.0 F 3.1
4:00 187.9 F 7.6
6:00 209.2 F 12.8
8:00 220.2 F 7.1
10:00 235.6 F 12.1
12:00 235.1 F 9.3
14:00 238.7 F 10.0
16:00 247.4 F 14.3
18:00 259.3 F 20.9
20:00 270.2 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.8
1:00 271.9 F 29.4
2:00 272.4 F 29.3
3:00 271.1 F 28.7
4:00 272.2 F 29.7

Min 271.1 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.7 F 0.0
5:00 231.0 F 0.0
10:00 207.7 F 0.0
15:00 184.2 F 0.0
20:00 178.9 F 0.0
25:00 175.7 F 0.0
30:00 172.9 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 20:49
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 00:59:49
CYCLE COMPLETE

09/12/2023 12:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____