

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2298

Name: RITTER2  
Sterilizer ID: V2313228  
04/25/2023 09:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.9 F 0.0  
2:00 184.1 F 3.6  
4:00 191.1 F 8.3  
6:00 213.0 F 9.3  
8:00 228.6 F 9.6  
10:00 233.4 F 8.7  
12:00 243.5 F 12.6  
14:00 245.3 F 13.2  
16:00 258.5 F 20.4  
18:00 270.1 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.9  
1:00 272.4 F 29.6  
2:00 271.5 F 28.9  
3:00 271.5 F 29.4  
4:00 272.9 F 29.8  
  
Min 271.0 F 28.8  
Max 273.1 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.0 F 0.0  
5:00 221.2 F 0.0  
10:00 234.3 F 0.0  
15:00 230.1 F 0.0  
20:00 227.4 F 0.0  
25:00 227.8 F 0.0  
30:00 227.3 F 0.0  
Drying Complete

FILLING: 1:36  
HEATING: 18:54  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:57:42  
CYCLE COMPLETE

04/25/2023 10:55 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_