

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1006

Name: RITTER 1  
Sterilizer ID: V2178198  
07/14/2021 06:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.9 F 0.0  
2:00 75.2 F 0.0  
4:00 77.3 F 0.0  
6:00 93.1 F 0.1  
8:00 119.8 F 0.1  
10:00 153.1 F 0.2  
12:00 180.5 F 0.2  
14:00 206.5 F 1.7  
16:00 218.0 F 4.4  
18:00 227.0 F 6.8  
20:00 233.6 F 10.1  
22:00 240.5 F 10.5  
24:00 240.6 F 11.1  
26:00 236.7 F 8.8  
28:00 241.2 F 10.3  
30:00 247.9 F 13.9  
32:00 259.4 F 20.1  
34:00 269.5 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 27.9  
1:00 272.6 F 28.6  
2:00 271.4 F 27.8  
3:00 272.4 F 28.7  
4:00 272.1 F 28.1  
  
Min 271.1 F 27.8  
Max 272.6 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.7 F 0.0  
5:00 233.1 F 0.0  
10:00 199.7 F 0.0  
15:00 184.2 F 0.0  
20:00 169.9 F 0.0  
25:00 163.8 F 0.0  
Dry Cycle Aborted

FILLING: 0:49  
HEATING: 35:02  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 29:12  
TOTAL CYCLE: 01:12:21

07/14/2021 07:55 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01