
Midmark M11 - v1.0.0
Total Cycles: 1763

Name: RITTER 1
Sterilizer ID: V2178198
03/29/2022 02:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.9 F 0.0
2:00 166.1 F 2.4
4:00 175.0 F 6.0
6:00 196.1 F 10.6
8:00 219.1 F 4.3
10:00 230.8 F 9.4
12:00 237.4 F 9.6
14:00 242.5 F 12.2
16:00 241.2 F 10.8
18:00 254.8 F 17.7
20:00 266.4 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.1
1:00 271.5 F 28.5
2:00 272.3 F 28.8
3:00 271.4 F 28.1
4:00 271.7 F 28.5

Min 271.0 F 27.9
Max 273.1 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 231.2 F 0.0
10:00 230.4 F 0.0
15:00 224.6 F 0.0
20:00 223.4 F 0.0
25:00 221.5 F 0.0
30:00 222.0 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:28
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:43
CYCLE COMPLETE

03/29/2022 03:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____