

-----  
Midmark M11 - v1.0.5  
Total Cycles: 261

Name: RITTER2  
Sterilizer ID: V2313228  
08/04/2021 09:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 101.0 F 0.0  
2:00 102.0 F 0.4  
4:00 109.6 F 1.7  
6:00 142.2 F 4.5  
8:00 169.0 F 7.6  
10:00 188.9 F 11.2  
12:00 201.3 F 4.2  
14:00 216.4 F 7.6  
16:00 226.7 F 11.8  
18:00 230.2 F 8.1  
20:00 242.1 F 13.2  
22:00 241.3 F 11.5  
24:00 252.5 F 17.2  
26:00 262.5 F 23.2  
28:00 271.9 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.2  
1:00 272.9 F 29.9  
2:00 271.7 F 29.1  
3:00 272.7 F 30.0  
4:00 272.3 F 29.5

Min 271.5 F 29.1  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.2 F 0.0  
5:00 228.1 F 0.0  
10:00 214.6 F 0.0  
15:00 211.9 F 0.0  
20:00 214.2 F 0.0  
25:00 216.2 F 0.0  
30:00 216.4 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 28:33  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:06:49  
CYCLE COMPLETE

08/04/2021 10:19 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01